

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112524\
 Data File : VY020429.D
 Acq On : 25 Nov 2024 15:02
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Nov 26 00:42:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 20 04:38:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	349	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	1713	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	4182	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	4658	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	776	193.535	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 387.080%#	
35) Dibromofluoromethane	7.646	113	502	45.702	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 91.400%	
50) Toluene-d8	10.115	98	2435	56.275	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 112.560%	
62) 4-Bromofluorobenzene	12.401	95	3821	274.701	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 549.400%#	
Target Compounds						
20) Methylene Chloride	4.622	84	464	113.990	ug/l	Qvalue # 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112524\
Data File : VY020429.D
Acq On : 25 Nov 2024 15:02
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Nov 26 00:42:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 20 04:38:24 2024
Response via : Initial Calibration

