

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112524\
 Data File : VY020438.D
 Acq On : 25 Nov 2024 18:31
 Operator : SY/MD
 Sample : P4954-03RE
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-06-112124RE

Quant Time: Nov 26 00:48:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 20 04:38:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	13111	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	20807	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	16571	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	4446	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	6413	42.574	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	85.140%
35) Dibromofluoromethane	7.634	113	6700	50.217	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.440%
50) Toluene-d8	10.109	98	23863	45.404	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	90.800%
62) 4-Bromofluorobenzene	12.407	95	6478	38.342	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	76.680%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112524\
Data File : VY020438.D
Acq On : 25 Nov 2024 18:31
Operator : SY/MD
Sample : P4954-03RE
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-06-112124RE

Quant Time: Nov 26 00:48:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 20 04:38:24 2024
Response via : Initial Calibration

