



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112524\
Data File : VY020422.D
Acq On : 25 Nov 2024 09:01
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
Title : SW846 8260
Last Update : Wed Nov 20 04:38:24 2024



AutoFind: Scans 1753, 1754, 1755; Background Corrected with Scan 1745

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.5	19509	PASS
75	95	30	60	54.8	52101	PASS
95	95	100	100	100.0	95093	PASS
96	95	5	9	6.4	6087	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	77.1	73309	PASS
175	174	5	9	8.0	5860	PASS
176	174	95	101	95.5	69989	PASS
177	176	5	9	6.4	4487	PASS