

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112723\
 Data File : VY016505.D
 Acq On : 27 Nov 2023 13:28
 Operator : SY/MD
 Sample : 05521-09
 Misc : 6.56g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SP-3-GRAB

Quant Time: Nov 28 00:16:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	98115	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	171536	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	167115	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	73723	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	56728	63.261	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 126.520%	
35) Dibromofluoromethane	7.734	113	54941	56.383	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 112.760%	
50) Toluene-d8	10.191	98	201609	51.021	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 102.040%	
62) 4-Bromofluorobenzene	12.495	95	68360	53.265	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 106.520%	
Target Compounds						
20) Methylene Chloride	4.735	84	17205	13.906	ug/l	Qvalue # 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112723\
Data File : VY016505.D
Acq On : 27 Nov 2023 13:28
Operator : SY/MD
Sample : 05521-09
Misc : 6.56g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SP-3-GRAB

Quant Time: Nov 28 00:16:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

