

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\YY112723\
 Data File : VY016499.D
 Acq On : 27 Nov 2023 10:06
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 28 00:14:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	84	0.00
2 T	Dichlorodifluoromethane	0.490	0.478	2.4	88	0.00
3 P	Chloromethane	0.573	0.519	9.4	81	0.00
4 C	Vinyl Chloride	0.617	0.593	3.9#	85	0.00
5 T	Bromomethane	0.405	0.397	2.0	87	0.00
6 T	Chloroethane	0.393	0.375	4.6	84	0.00
7 T	Trichlorofluoromethane	0.897	0.949	-5.8	90	0.00
8 T	Diethyl Ether	0.268	0.281	-4.9	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.540	0.542	-0.4	88	0.00
10 T	Methyl Iodide	0.584	0.663	-13.5	94	0.00
11 T	Tert butyl alcohol	0.036	0.033	8.3	81	0.00
12 CM	1,1-Dichloroethene	0.505	0.528	-4.6#	91	0.00
13 T	Acrolein	0.043	0.044	-2.3	85	0.00
14 T	Allyl chloride	0.717	0.623	13.1	73	0.00
15 T	Acrylonitrile	0.106	0.106	0.0	82	0.00
16 T	Acetone	0.076	0.095	-25.0#	111	0.00
17 T	Carbon Disulfide	1.525	1.524	0.1	86	0.00
18 T	Methyl Acetate	0.402	0.389	3.2	81	0.00
19 T	Methyl tert-butyl Ether	1.223	1.264	-3.4	86	0.00
20 T	Methylene Chloride	0.610	0.576	5.6	88	0.00
21 T	trans-1,2-Dichloroethene	0.582	0.600	-3.1	90	0.00
22 T	Diisopropyl ether	1.545	1.373	11.1	74	0.00
23 T	Vinyl Acetate	0.749	0.662	11.6	72	0.00
24 P	1,1-Dichloroethane	0.953	0.916	3.9	82	0.00
25 T	2-Butanone	0.132	0.132	0.0	84	0.00
26 T	2,2-Dichloropropane	0.868	0.883	-1.7	87	0.00
27 T	cis-1,2-Dichloroethene	0.627	0.662	-5.6	90	0.00
28 T	Bromochloromethane	0.353	0.322	8.8	78	0.00
29 T	Tetrahydrofuran	0.088	0.080	9.1	74	0.00
30 C	Chloroform	0.992	1.028	-3.6#	89	0.00
31 T	Cyclohexane	0.900	0.767	14.8	76	0.00
32 T	1,1,1-Trichloroethane	0.910	0.969	-6.5	91	0.00
33 S	1,2-Dichloroethane-d4	0.457	0.420	8.1	79	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	82	0.00
35 S	Dibromofluoromethane	0.284	0.310	-9.2	89	0.00
36 T	1,1-Dichloropropene	0.493	0.500	-1.4	83	0.00
37 T	Ethyl Acetate	0.199	0.181	9.0	73	0.00
38 T	Carbon Tetrachloride	0.512	0.586	-14.5	93	0.00
39 T	Methylcyclohexane	0.615	0.630	-2.4	83	0.00
40 TM	Benzene	1.386	1.470	-6.1	86	0.00
41 T	Methacrylonitrile	0.100	0.102	-2.0	85	0.00
42 TM	1,2-Dichloroethane	0.359	0.385	-7.2	86	0.00
43 T	Isopropyl Acetate	0.385	0.356	7.5	73	0.00
44 TM	Trichloroethene	0.398	0.447	-12.3	92	0.00
45 C	1,2-Dichloropropane	0.333	0.337	-1.2#	80	0.00
46 T	Dibromomethane	0.180	0.200	-11.1	88	0.00
47 T	Bromodichloromethane	0.464	0.506	-9.1	88	0.00
48 T	Methyl methacrylate	0.213	0.203	4.7	74	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112723\
 Data File : VY016499.D
 Acq On : 27 Nov 2023 10:06
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 28 00:14:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	85	0.00
50 S	Toluene-d8	1.152	1.122	2.6	80	0.00
51 T	4-Methyl-2-Pentanone	0.195	0.184	5.6	74	0.00
52 CM	Toluene	0.876	0.948	-8.2#	87	0.00
53 T	t-1,3-Dichloropropene	0.449	0.490	-9.1	86	0.00
54 T	cis-1,3-Dichloropropene	0.537	0.580	-8.0	86	0.00
55 T	1,1,2-Trichloroethane	0.249	0.283	-13.7	90	0.00
56 T	Ethyl methacrylate	0.235	0.256	-8.9	83	0.00
57 T	1,3-Dichloropropane	0.424	0.468	-10.4	87	0.00
58 T	2-Chloroethyl Vinyl ether	0.159	0.161	-1.3	74	0.00
59 T	2-Hexanone	0.145	0.137	5.5	78	0.00
60 T	Dibromochloromethane	0.311	0.372	-19.6	94	0.00
61 T	1,2-Dibromoethane	0.237	0.271	-14.3	90	0.00
62 S	4-Bromofluorobenzene	0.374	0.385	-2.9	85	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	83	0.00
64 T	Tetrachloroethene	0.449	0.513	-14.3	94	0.00
65 PM	Chlorobenzene	1.068	1.171	-9.6	90	0.00
66 T	1,1,1,2-Tetrachloroethane	0.394	0.450	-14.2	93	0.00
67 C	Ethyl Benzene	1.920	2.036	-6.0#	86	0.00
68 T	m/p-Xylenes	0.728	0.793	-8.9	88	0.00
69 T	o-Xylene	0.676	0.751	-11.1	88	0.00
70 T	Styrene	0.968	1.088	-12.4	90	0.00
71 P	Bromoform	0.204	0.251	-23.0	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
73 T	Isopropylbenzene	4.049	4.321	-6.7	89	0.00
74 T	N-ethyl acetate	0.813	0.710	12.7	69	0.00
75 P	1,1,2,2-Tetrachloroethane	0.650	0.689	-6.0	88	0.00
76 T	1,2,3-Trichloropropane	0.480	0.521	-8.5	96	0.00
77 T	Bromobenzene	0.915	1.018	-11.3	92	0.00
78 T	n-propylbenzene	4.785	4.926	-2.9	86	0.00
79 T	2-Chlorotoluene	2.637	2.771	-5.1	87	0.00
80 T	1,3,5-Trimethylbenzene	3.264	3.494	-7.0	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.221	0.238	-7.7	89	0.00
82 T	4-Chlorotoluene	2.725	2.838	-4.1	87	0.00
83 T	tert-Butylbenzene	2.875	3.071	-6.8	88	0.00
84 T	1,2,4-Trimethylbenzene	3.214	3.470	-8.0	89	0.00
85 T	sec-Butylbenzene	4.108	4.280	-4.2	87	0.00
86 T	p-Isopropyltoluene	3.537	3.765	-6.4	88	0.00
87 T	1,3-Dichlorobenzene	1.761	1.905	-8.2	92	0.00
88 T	1,4-Dichlorobenzene	1.721	1.879	-9.2	92	0.00
89 T	n-Butylbenzene	3.267	3.341	-2.3	84	0.00
90 T	Hexachloroethane	0.682	0.735	-7.8	91	0.00
91 T	1,2-Dichlorobenzene	1.509	1.660	-10.0	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.103	0.104	-1.0	88	0.00
93 T	1,2,4-Trichlorobenzene	0.881	0.953	-8.2	90	0.00
94 T	Hexachlorobutadiene	0.504	0.579	-14.9	97	0.00
95 T	Naphthalene	1.575	1.698	-7.8	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112723\
Data File : VY016499.D
Acq On : 27 Nov 2023 10:06
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Nov 28 00:14:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.728	0.775	-6.5	92	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6