

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112823\
 Data File : VY016516.D
 Acq On : 28 Nov 2023 13:26
 Operator : SY/MD
 Sample : 05521-12
 Misc : 5.60g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SP-4-GRAB

Quant Time: Nov 29 03:07:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.813	168	107406	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.716	114	186865	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.508	117	173698	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.447	152	71558	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.167	65	56673	57.732	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.460%
35) Dibromofluoromethane	7.746	113	59053	55.631	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.260%
50) Toluene-d8	10.203	98	215973	50.172	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.340%
62) 4-Bromofluorobenzene	12.502	95	67902	48.568	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	97.140%
Target Compounds						Qvalue
20) Methylene Chloride	4.753	84	24229	18.961	ug/l	# 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112823\
Data File : VY016516.D
Acq On : 28 Nov 2023 13:26
Operator : SY/MD
Sample : 05521-12
Misc : 5.60g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SP-4-GRAB

Quant Time: Nov 29 03:07:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

