



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112823\
 Data File : VY016510.D
 Acq On : 28 Nov 2023 09:47
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Title : SW846 8260
 Last Update : Thu Nov 09 00:45:52 2023



AutoFind: Scans 1751, 1752, 1753; Background Corrected with Scan 1743

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.7	12091	PASS
75	95	30	60	48.7	37480	PASS
95	95	100	100	100.0	76981	PASS
96	95	5	9	6.5	5019	PASS
173	174	0.00	2	1.0	701	PASS
174	95	50	100	87.5	67395	PASS
175	174	5	9	7.4	4964	PASS
176	174	95	101	97.1	65467	PASS
177	176	5	9	7.1	4638	PASS