

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112923\
 Data File : VY016532.D
 Acq On : 29 Nov 2023 12:05
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Nov 30 03:10:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 30 02:51:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.819	168	105945	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.722	114	173967	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.514	117	158929	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.446	152	63100	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.173	65	48492	57.935	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.860%
35) Dibromofluoromethane	7.746	113	51559	50.497	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.000%
50) Toluene-d8	10.203	98	200800	50.706	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.420%
62) 4-Bromofluorobenzene	12.508	95	60746	48.898	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	97.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112923\
Data File : VY016532.D
Acq On : 29 Nov 2023 12:05
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Nov 30 03:10:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112923S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 30 02:51:41 2023
Response via : Initial Calibration

