



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112923\
Data File : VY016525.D
Acq On : 29 Nov 2023 07:53
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112923S.M
Title : SW846 8260
Last Update : Thu Nov 30 03:28:04 2023



AutoFind: Scans 1753, 1754, 1755; Background Corrected with Scan 1745

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.2	10820	PASS
75	95	30	60	48.6	34560	PASS
95	95	100	100	100.0	71104	PASS
96	95	5	9	6.7	4750	PASS
173	174	0.00	2	0.9	583	PASS
174	95	50	100	89.4	63571	PASS
175	174	5	9	7.4	4679	PASS
176	174	95	101	97.2	61773	PASS
177	176	5	9	6.6	4053	PASS