

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY113022\
 Data File : VY011636.D
 Acq On : 30 Nov 2022 13:18
 Operator : KP/MD
 Sample : N5809-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HOWE-DRUM-A-1129

Quant Time: Dec 01 05:22:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 19 22:37:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	1359	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.697	114	2409	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	1915	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	312	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	1254	90.875	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	181.740%#
35) Dibromofluoromethane	7.728	113	463	31.221	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	62.440%
50) Toluene-d8	10.185	98	2573	47.746	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.500%
62) 4-Bromofluorobenzene	12.495	95	491	24.643	ug/l	0.01
Spiked Amount	50.000	Range	25 - 144	Recovery	=	49.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY113022\
Data File : VY011636.D
Acq On : 30 Nov 2022 13:18
Operator : KP/MD
Sample : N5809-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HOWE-DRUM-A-1129

Quant Time: Dec 01 05:22:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 19 22:37:14 2022
Response via : Initial Calibration

