

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY113022\
 Data File : VY011638.D
 Acq On : 30 Nov 2022 14:05
 Operator : KP/MD
 Sample : N5809-05
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HOWE-DRUM-C-1129

Quant Time: Dec 01 05:23:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 19 22:37:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	843	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.697	114	1432	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	1194	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	254	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	1090	127.340	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	254.680%#
35) Dibromofluoromethane	7.710	113	418	47.418	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.840%
50) Toluene-d8	10.179	98	1871	58.407	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	116.820%
62) 4-Bromofluorobenzene	12.489	95	334	28.200	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	56.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY113022\
Data File : VY011638.D
Acq On : 30 Nov 2022 14:05
Operator : KP/MD
Sample : N5809-05
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HOWE-DRUM-C-1129

Quant Time: Dec 01 05:23:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 19 22:37:14 2022
Response via : Initial Calibration

