

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY113023\
 Data File : VY016546.D
 Acq On : 30 Nov 2023 11:40
 Operator : SY/MD
 Sample : 05576-01
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-04-112923

Quant Time: Dec 01 00:58:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 30 03:28:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.807	168	103141	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.709	114	174988	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.508	117	165184	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	74808	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	55998	68.721	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	137.440%
35) Dibromofluoromethane	7.740	113	54910	53.465	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.940%
50) Toluene-d8	10.197	98	201926	50.693	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.380%
62) 4-Bromofluorobenzene	12.495	95	67293	53.852	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	107.700%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY113023\
 Data File : VY016546.D
 Acq On : 30 Nov 2023 11:40
 Operator : SY/MD
 Sample : 05576-01
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-04-112923

Quant Time: Dec 01 00:58:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 30 03:28:04 2023
 Response via : Initial Calibration

