

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120123\
 Data File : VY016553.D
 Acq On : 01 Dec 2023 09:57
 Operator : SY/MD
 Sample : VY1201SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1201SBL01

Quant Time: Dec 01 23:35:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 30 03:28:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	103925	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.710	114	172389	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.508	117	160159	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	67719	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	50757	61.819	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.640%
35) Dibromofluoromethane	7.740	113	53430	52.809	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.620%
50) Toluene-d8	10.197	98	197223	50.258	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.520%
62) 4-Bromofluorobenzene	12.495	95	63597	51.662	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	103.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120123\
Data File : VY016553.D
Acq On : 01 Dec 2023 09:57
Operator : SY/MD
Sample : VY1201SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1201SBL01

Quant Time: Dec 01 23:35:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112923S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 30 03:28:04 2023
Response via : Initial Calibration

