

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120123\
 Data File : VY016556.D
 Acq On : 01 Dec 2023 12:07
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Dec 01 23:36:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 30 03:28:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.813	168	103031	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.709	114	167715	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.508	117	146093	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.446	152	55886	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.167	65	42401	52.090	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.180%
35) Dibromofluoromethane	7.740	113	48390	49.160	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.320%
50) Toluene-d8	10.197	98	190517	49.903	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.800%
62) 4-Bromofluorobenzene	12.501	95	54418	45.437	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	90.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120123\
Data File : VY016556.D
Acq On : 01 Dec 2023 12:07
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Dec 01 23:36:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112923S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 30 03:28:04 2023
Response via : Initial Calibration

