

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120123\
 Data File : VY016558.D
 Acq On : 01 Dec 2023 13:46
 Operator : SY/MD
 Sample : 05586-01
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VAULT-1

Quant Time: Dec 01 23:37:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 30 03:28:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.820	168	104773	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.716	114	178326	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.508	117	173807	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	80168	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.167	65	56281	67.993	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.980%
35) Dibromofluoromethane	7.746	113	54396	51.973	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.940%
50) Toluene-d8	10.203	98	207880	51.210	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.420%
62) 4-Bromofluorobenzene	12.496	95	71947	56.499	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	113.000%
Target Compounds						Qvalue
20) Methylene Chloride	4.753	84	8737	6.698	ug/l	# 73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120123\
Data File : VY016558.D
Acq On : 01 Dec 2023 13:46
Operator : SY/MD
Sample : 05586-01
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VAULT-1

Quant Time: Dec 01 23:37:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112923S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 30 03:28:04 2023
Response via : Initial Calibration

