

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120123\
 Data File : VY016560.D
 Acq On : 01 Dec 2023 14:33
 Operator : SY/MD
 Sample : 05590-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SH-SOIL-A

Quant Time: Dec 01 23:38:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 30 03:28:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.813	168	3481	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.716	114	5356	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.514	117	1976	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	247	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.167	65	1332	48.434	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.860%
35) Dibromofluoromethane	7.740	113	1451	46.159	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.320%
50) Toluene-d8	10.203	98	4829	39.608	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	79.220%
62) 4-Bromofluorobenzene	12.501	95	323	8.445	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	16.900%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120123\
Data File : VY016560.D
Acq On : 01 Dec 2023 14:33
Operator : SY/MD
Sample : 05590-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SH-SOIL-A

Quant Time: Dec 01 23:38:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112923S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 30 03:28:04 2023
Response via : Initial Calibration

