

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120123\
 Data File : VY016566.D
 Acq On : 01 Dec 2023 17:23
 Operator : SY/MD
 Sample : 05602-03
 Misc : 4.12g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SPRING-VALLEY-TP-2

Quant Time: Dec 01 23:40:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 30 03:28:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	105402	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.703	114	179665	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.502	117	173255	50.000	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.434	152	81753	50.000	ug/l	-0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	58606	70.379	ug/l	-0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	140.760%
35) Dibromofluoromethane	7.734	113	55462	52.597	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.200%
50) Toluene-d8	10.191	98	204555	50.016	ug/l	-0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.040%
62) 4-Bromofluorobenzene	12.489	95	72621	56.603	ug/l	-0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	113.200%

Target Compounds			Qvalue			
16) Acetone	3.979	43	2341	18.474	ug/l	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120123\
Data File : VY016566.D
Acq On : 01 Dec 2023 17:23
Operator : SY/MD
Sample : 05602-03
Misc : 4.12g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SPRING-VALLEY-TP-2

Quant Time: Dec 01 23:40:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112923S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 30 03:28:04 2023
Response via : Initial Calibration

