

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120123\
 Data File : VY016567.D
 Acq On : 01 Dec 2023 17:46
 Operator : SY/MD
 Sample : 05602-07
 Misc : 6.48g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ROUTE-4-TP-4

Quant Time: Dec 01 23:41:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 30 03:28:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	98477	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.703	114	165076	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.502	117	161887	50.000	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.434	152	78923	50.000	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	54967	70.651	ug/l	-0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	141.300%
35) Dibromofluoromethane	7.734	113	52915	54.617	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.240%
50) Toluene-d8	10.191	98	191044	50.841	ug/l	-0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.680%
62) 4-Bromofluorobenzene	12.495	95	67846	57.555	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	115.100%
Target Compounds						
16) Acetone	3.966	43	6129	42.518	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120123\
Data File : VY016567.D
Acq On : 01 Dec 2023 17:46
Operator : SY/MD
Sample : 05602-07
Misc : 6.48g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ROUTE-4-TP-4

Quant Time: Dec 01 23:41:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112923S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 30 03:28:04 2023
Response via : Initial Calibration

