

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120123\
 Data File : VY016568.D
 Acq On : 01 Dec 2023 18:09
 Operator : SY/MD
 Sample : 05602-11
 Misc : 5.23g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ROUTE-4-TP-6

Quant Time: Dec 01 23:41:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 30 03:28:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	107274	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.703	114	177055	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.502	117	171059	50.000	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.440	152	79329	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	57447	67.783	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 135.560%	
35) Dibromofluoromethane	7.734	113	54610	52.552	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 105.100%	
50) Toluene-d8	10.191	98	203828	50.573	ug/l	-0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.140%	
62) 4-Bromofluorobenzene	12.495	95	71922	56.884	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 113.760%	
Target Compounds						Qvalue
16) Acetone	3.967	43	29667	178.230	ug/l #	86
25) 2-Butanone	7.009	43	12176	53.249	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120123\
Data File : VY016568.D
Acq On : 01 Dec 2023 18:09
Operator : SY/MD
Sample : 05602-11
Misc : 5.23g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ROUTE-4-TP-6

Quant Time: Dec 01 23:41:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112923S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 30 03:28:04 2023
Response via : Initial Calibration

