

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120222\
 Data File : VY011675.D
 Acq On : 02 Dec 2022 12:05
 Operator : KP/MD
 Sample : VY1202SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1202SBL01

Quant Time: Dec 02 23:50:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 19 22:37:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	168009	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	279950	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	254079	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	118129	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	103668	60.768	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.540%
35) Dibromofluoromethane	7.722	113	97448	56.546	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.100%
50) Toluene-d8	10.179	98	309107	49.358	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.720%
62) 4-Bromofluorobenzene	12.483	95	113704	49.107	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120222\
Data File : VY011675.D
Acq On : 02 Dec 2022 12:05
Operator : KP/MD
Sample : VY1202SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1202SBL01

Quant Time: Dec 02 23:50:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 19 22:37:14 2022
Response via : Initial Calibration

