

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120222\
 Data File : VY011682.D
 Acq On : 02 Dec 2022 14:48
 Operator : KP/MD
 Sample : N5845-01
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-03-120122

Quant Time: Dec 02 23:52:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 19 22:37:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	1832	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	3645	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	4004	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	1116	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	2356	126.653	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	253.300%#
35) Dibromofluoromethane	7.722	113	853	38.015	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	76.040%
50) Toluene-d8	10.185	98	4485	55.004	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.000%
62) 4-Bromofluorobenzene	12.483	95	1494	49.557	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120222\
Data File : VY011682.D
Acq On : 02 Dec 2022 14:48
Operator : KP/MD
Sample : N5845-01
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-03-120122

Quant Time: Dec 02 23:52:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 19 22:37:14 2022
Response via : Initial Calibration

