

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120222\
 Data File : VY011684.D
 Acq On : 02 Dec 2022 15:35
 Operator : KP/MD
 Sample : N5868-01
 Misc : 3.60g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-0

Quant Time: Dec 02 23:53:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 19 22:37:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	185001	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	340060	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	338661	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	151250	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	127641	67.949	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.900%
35) Dibromofluoromethane	7.716	113	110893	52.973	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.940%
50) Toluene-d8	10.179	98	419504	55.146	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.300%
62) 4-Bromofluorobenzene	12.483	95	137115	48.750	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120222\
Data File : VY011684.D
Acq On : 02 Dec 2022 15:35
Operator : KP/MD
Sample : N5868-01
Misc : 3.60g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-0

Quant Time: Dec 02 23:53:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 19 22:37:14 2022
Response via : Initial Calibration

