

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120224\
 Data File : VY020480.D
 Acq On : 02 Dec 2024 12:54
 Operator : SY/MD
 Sample : VY1202SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1202SBL01

Quant Time: Dec 02 14:50:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120224S.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 02 14:47:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.726	168	165618	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.628	114	314207	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.426	117	283582	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.359	152	97585	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.079	65	96612	57.358	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.720%
35) Dibromofluoromethane	7.652	113	98098	49.592	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.180%
50) Toluene-d8	10.115	98	380303	50.490	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.980%
62) 4-Bromofluorobenzene	12.414	95	117100	44.451	ug/l	0.01
Spiked Amount	50.000	Range	29 - 146	Recovery	=	88.900%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSV0A_Y\Data\VY120224\
Data File : VY020480.D
Acq On : 02 Dec 2024 12:54
Operator : SY/MD
Sample : VY1202SBL01
Misc : 5.00g/5.0mL/MSV0A_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSV0A_Y
ClientSampleId :
VY1202SBL01

Quant Time: Dec 02 14:50:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSV0A_Y\methods\82Y120224S.M
Quant Title : SW846 8260
QLast Update : Mon Dec 02 14:47:59 2024
Response via : Initial Calibration

