

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120224\
 Data File : VY020484.D
 Acq On : 02 Dec 2024 14:47
 Operator : SY/MD
 Sample : P5005-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 STOCK-PILE

Quant Time: Dec 02 21:56:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120224S.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 02 14:47:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.720	168	688	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.628	114	895	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.426	117	312	50.000	ug/l	# 0.01
72) 1,4-Dichlorobenzene-d4	13.365	152	90	50.000	ug/l	# 0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	493	70.458	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 140.920%	
35) Dibromofluoromethane	7.640	113	424	75.251	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 150.500%#	
50) Toluene-d8	10.128	98	836	38.965	ug/l	0.02
Spiked Amount	50.000	Range	58 - 134	Recovery	= 77.940%	
62) 4-Bromofluorobenzene	12.420	95	94	12.527	ug/l	0.02
Spiked Amount	50.000	Range	29 - 146	Recovery	= 25.060%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSV0A_Y\Data\VY120224\
Data File : VY020484.D
Acq On : 02 Dec 2024 14:47
Operator : SY/MD
Sample : P5005-01
Misc : 5.08g/5.0mL/MSV0A_Y/SOIL/B
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSV0A_Y
ClientSampleId :
STOCK-PILE

Quant Time: Dec 02 21:56:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSV0A_Y\methods\82Y120224S.M
Quant Title : SW846 8260
QLast Update : Mon Dec 02 14:47:59 2024
Response via : Initial Calibration

