

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120224\
 Data File : VY020486.D
 Acq On : 02 Dec 2024 15:34
 Operator : SY/MD
 Sample : P5046-01
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HD-01-11272024

Quant Time: Dec 02 21:57:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120224S.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 02 14:47:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	15239	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	23447	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	15711	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	3154	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	6029	38.901	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	77.800%
35) Dibromofluoromethane	7.646	113	7031	47.632	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.260%
50) Toluene-d8	10.115	98	24970	44.425	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	88.840%
62) 4-Bromofluorobenzene	12.414	95	5355	27.240	ug/l	0.01
Spiked Amount	50.000	Range	29 - 146	Recovery	=	54.480%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120224\
Data File : VY020486.D
Acq On : 02 Dec 2024 15:34
Operator : SY/MD
Sample : P5046-01
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HD-01-11272024

Quant Time: Dec 02 21:57:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120224S.M
Quant Title : SW846 8260
QLast Update : Mon Dec 02 14:47:59 2024
Response via : Initial Calibration

