

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120224\
 Data File : VY020488.D
 Acq On : 02 Dec 2024 16:21
 Operator : SY/MD
 Sample : P5026-04
 Misc : 6.51g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SOIL-1-HAM-GRAB

Quant Time: Dec 02 21:58:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120224S.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 02 14:47:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	162960	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	318573	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.426	117	289732	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.352	152	105689	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	94036	56.739	ug/l	0.01
Spiked Amount 50.000	Range 50 - 163		Recovery = 113.480%			
35) Dibromofluoromethane	7.646	113	55593	27.719	ug/l	0.01
Spiked Amount 50.000	Range 54 - 147		Recovery = 55.440%			
50) Toluene-d8	10.115	98	384370	50.331	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 100.660%			
62) 4-Bromofluorobenzene	12.414	95	119499	44.740	ug/l	0.01
Spiked Amount 50.000	Range 29 - 146		Recovery = 89.480%			
Target Compounds						
16) Acetone	3.897	43	5671	13.761	ug/l #	88
17) Carbon Disulfide	4.122	76	26927	5.562	ug/l	96
52) Toluene	10.176	92	10761	1.727	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120224\
Data File : VY020488.D
Acq On : 02 Dec 2024 16:21
Operator : SY/MD
Sample : P5026-04
Misc : 6.51g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SOIL-1-HAM-GRAB

Quant Time: Dec 02 21:58:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120224S.M
Quant Title : SW846 8260
QLast Update : Mon Dec 02 14:47:59 2024
Response via : Initial Calibration

