

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv120419\
 Data File : VY000869.D
 Acq On : 04 Dec 2019 18:27
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 05 08:27:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y112719S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 27 12:40:47 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	86	0.00
2 T	Dichlorodifluoromethane	0.538	0.445	17.3	71	0.00
3 P	Chloromethane	0.663	0.606	8.6	77	0.00
4 C	Vinyl Chloride	0.654	0.599	8.4#	77	0.00
5 T	Bromomethane	0.394	0.378	4.1	81	0.00
6 T	Chloroethane	0.401	0.375	6.5	77	0.00
7 T	Trichlorofluoromethane	0.896	0.819	8.6	77	0.00
8 T	Diethyl Ether	0.337	0.339	-0.6	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.553	0.508	8.1	77	0.00
10 T	Methyl Iodide	0.697	0.650	6.7	76	0.00
11 T	Tert butyl alcohol	0.053	0.052	1.9	76	0.00
12 CM	1,1-Dichloroethene	0.548	0.512	6.6#	79	0.00
13 T	Acrolein	0.053	0.052	1.9	86	0.00
14 T	Allyl chloride	0.940	0.929	1.2	80	0.00
15 T	Acrylonitrile	0.174	0.177	-1.7	82	0.00
16 T	Acetone	0.122	0.121	0.8	83	0.00
17 T	Carbon Disulfide	1.773	1.602	9.6	75	0.00
18 T	Methyl Acetate	0.510	0.500	2.0	82	0.00
19 T	Methyl tert-butyl Ether	1.464	1.507	-2.9	81	0.00
20 T	Methylene Chloride	0.642	0.626	2.5	83	0.00
21 T	trans-1,2-Dichloroethene	0.623	0.592	5.0	79	0.00
22 T	Diisopropyl ether	1.932	2.000	-3.5	83	0.00
23 T	Vinyl Acetate	1.175	1.229	-4.6	81	0.00
24 P	1,1-Dichloroethane	1.065	1.054	1.0	81	0.00
25 T	2-Butanone	0.225	0.224	0.4	75	0.00
26 T	2,2-Dichloropropane	0.896	0.840	6.3	78	0.00
27 T	cis-1,2-Dichloroethene	0.674	0.674	0.0	82	0.00
28 T	Bromochloromethane	0.360	0.460	-27.8#	96	0.00
29 T	Tetrahydrofuran	0.149	0.156	-4.7	81	0.00
30 C	Chloroform	1.068	1.013	5.1#	81	0.00
31 T	Cyclohexane	1.128	1.030	8.7	77	0.00
32 T	1,1,1-Trichloroethane	0.882	0.845	4.2	80	0.00
33 S	1,2-Dichloroethane-d4	0.531	0.543	-2.3	88	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.291	0.287	1.4	90	0.00
36 T	1,1-Dichloropropene	0.497	0.446	10.3	79	0.00
37 T	Ethyl Acetate	0.287	0.272	5.2	79	0.00
38 T	Carbon Tetrachloride	0.443	0.404	8.8	79	0.00
39 T	Methylcyclohexane	0.659	0.587	10.9	78	0.00
40 TM	Benzene	1.448	1.380	4.7	82	0.00
41 T	Methacrylonitrile	0.151	0.154	-2.0	89	0.00
42 TM	1,2-Dichloroethane	0.381	0.380	0.3	85	0.00
43 T	Isopropyl Acetate	0.541	0.518	4.3	79	0.00
44 TM	Trichloroethene	0.372	0.350	5.9	81	0.00
45 C	1,2-Dichloropropane	0.364	0.357	1.9#	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv120419\
 Data File : VY000869.D
 Acq On : 04 Dec 2019 18:27
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 05 08:27:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y112719S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 27 12:40:47 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.190	0.189	0.5	83	0.00
47 T	Bromodichloromethane	0.446	0.439	1.6	82	0.00
48 T	Methyl methacrylate	0.238	0.228	4.2	75	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	79	0.00
50 S	Toluene-d8	1.165	1.162	0.3	87	0.00
51 T	4-Methyl-2-Pentanone	0.280	0.279	0.4	81	0.00
52 CM	Toluene	0.897	0.862	3.9#	81	0.00
53 T	t-1,3-Dichloropropene	0.480	0.485	-1.0	82	0.00
54 T	cis-1,3-Dichloropropene	0.567	0.561	1.1	83	0.00
55 T	1,1,2-Trichloroethane	0.281	0.274	2.5	82	0.00
56 T	Ethyl methacrylate	0.402	0.404	-0.5	80	0.00
57 T	1,3-Dichloropropane	0.489	0.493	-0.8	84	0.00
58 T	2-Chloroethyl Vinyl ether	0.162	0.150	7.4	82	0.00
59 T	2-Hexanone	0.204	0.190	6.9	70	0.00
60 T	Dibromochloromethane	0.302	0.304	-0.7	83	0.00
61 T	1,2-Dibromoethane	0.265	0.265	0.0	82	0.00
62 S	4-Bromofluorobenzene	0.438	0.412	5.9	88	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.434	0.404	6.9	83	0.00
65 PM	Chlorobenzene	1.039	0.969	6.7	81	0.00
66 T	1,1,1,2-Tetrachloroethane	0.350	0.339	3.1	83	0.00
67 C	Ethyl Benzene	1.936	1.811	6.5#	82	0.00
68 T	m/p-Xylenes	0.727	0.676	7.0	81	0.00
69 T	o-Xylene	0.677	0.648	4.3	83	0.00
70 T	Styrene	1.160	1.127	2.8	83	0.00
71 P	Bromoform	0.198	0.195	1.5	81	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	3.981	3.725	6.4	81	0.00
74 T	N-amyl acetate	1.108	1.078	2.7	80	0.00
75 P	1,1,2,2-Tetrachloroethane	0.784	0.764	2.6	80	0.00
76 T	1,2,3-Trichloropropane	0.566	0.559	1.2	85	0.00
77 T	Bromobenzene	0.863	0.827	4.2	82	0.00
78 T	n-propylbenzene	4.861	4.560	6.2	81	0.00
79 T	2-Chlorotoluene	2.685	2.517	6.3	82	0.00
80 T	1,3,5-Trimethylbenzene	3.296	3.113	5.6	82	0.00
81 T	trans-1,4-Dichloro-2-butene	0.288	0.281	2.4	76	0.00
82 T	4-Chlorotoluene	2.767	2.613	5.6	83	0.00
83 T	tert-Butylbenzene	2.771	2.597	6.3	81	0.00
84 T	1,2,4-Trimethylbenzene	3.298	3.086	6.4	81	0.00
85 T	sec-Butylbenzene	4.030	3.787	6.0	82	0.00
86 T	p-Isopropyltoluene	3.506	3.303	5.8	81	0.00
87 T	1,3-Dichlorobenzene	1.693	1.613	4.7	83	0.00
88 T	1,4-Dichlorobenzene	1.676	1.597	4.7	82	0.00
89 T	n-Butylbenzene	3.544	3.329	6.1	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv120419\
Data File : VY000869.D
Acq On : 04 Dec 2019 18:27
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Dec 05 08:27:36 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y112719S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 27 12:40:47 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.646	0.624	3.4	81	0.00
91 T	1,2-Dichlorobenzene	1.523	1.487	2.4	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.131	0.128	2.3	80	0.00
93 T	1,2,4-Trichlorobenzene	1.030	0.988	4.1	82	0.00
94 T	Hexachlorobutadiene	0.511	0.467	8.6	80	0.00
95 T	Naphthalene	2.393	2.345	2.0	82	0.00
96 T	1,2,3-Trichlorobenzene	0.920	0.883	4.0	83	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6