

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120423\
 Data File : VY016584.D
 Acq On : 04 Dec 2023 18:07
 Operator : SY/MD
 Sample : 05538-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-02-112223

Quant Time: Dec 05 01:31:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 05 01:24:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	7619	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.709	114	10964	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.501	117	10199	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	3513	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.160	65	4416	62.709	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	125.420%
35) Dibromofluoromethane	7.740	113	2543	38.541	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	77.080%
50) Toluene-d8	10.197	98	12664	50.741	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.480%
62) 4-Bromofluorobenzene	12.501	95	3746	45.641	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	91.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120423\
Data File : VY016584.D
Acq On : 04 Dec 2023 18:07
Operator : SY/MD
Sample : 05538-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-02-112223

Quant Time: Dec 05 01:31:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120423S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 05 01:24:25 2023
Response via : Initial Calibration

