

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY120519\
 Data File : VY000880.D
 Acq On : 05 Dec 2019 14:11
 Operator : SY/MD
 Sample : K6147-04
 Misc : 5.19G/5ML/MSVOA Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P001-WC004

Quant Time: Dec 06 04:42:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y112719S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 27 12:40:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.80	168	212931	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.70	114	407886	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.49	117	365841	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.42	152	158294	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.15	65	132380	58.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.02%	
35) Dibromofluoromethane	7.73	113	117880	49.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.18%	
50) Toluene-d8	10.18	98	485687	51.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.20%	
62) 4-Bromofluorobenzene	12.48	95	160661	45.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.02%	
Target Compounds						
44) Trichloroethene	8.94	130	8924	2.939	ug/l	97
64) Tetrachloroethene	10.72	164	104030	32.729	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY120519\
Data File : VY000880.D
Acq On : 05 Dec 2019 14:11
Operator : SY/MD
Sample : K6147-04
Misc : 5.19G/5ML/MSVOA Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P001-WC004

Quant Time: Dec 06 04:42:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y112719S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 27 12:40:47 2019
Response via : Initial Calibration

