

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120522\
 Data File : VY011696.D
 Acq On : 05 Dec 2022 15:05
 Operator : KP/MD
 Sample : N5845-01RE
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-03-120122RE

Quant Time: Dec 06 00:36:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 19 22:37:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	1659	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	3065	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	3072	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	558	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	1831	108.694	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 217.380%#	
35) Dibromofluoromethane	7.722	113	581	30.793	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 61.580%	
50) Toluene-d8	10.185	98	3790	55.277	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 110.560%	
62) 4-Bromofluorobenzene	12.477	95	1005	39.645	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 79.280%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120522\
Data File : VY011696.D
Acq On : 05 Dec 2022 15:05
Operator : KP/MD
Sample : N5845-01RE
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-03-120122RE

Quant Time: Dec 06 00:36:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 19 22:37:14 2022
Response via : Initial Calibration

