

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120522\
 Data File : VY011699.D
 Acq On : 05 Dec 2022 16:16
 Operator : KP/MD
 Sample : N5872-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-02-120222

Quant Time: Dec 06 00:37:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 19 22:37:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	62414	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	113579	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	120324	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	54808	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	53174	83.904	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	167.800%#
35) Dibromofluoromethane	7.716	113	42688	61.054	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	122.100%
50) Toluene-d8	10.179	98	139303	54.827	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.660%
62) 4-Bromofluorobenzene	12.483	95	47023	50.056	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120522\
Data File : VY011699.D
Acq On : 05 Dec 2022 16:16
Operator : KP/MD
Sample : N5872-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-02-120222

Quant Time: Dec 06 00:37:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 19 22:37:14 2022
Response via : Initial Calibration

