

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120523\
 Data File : VY016599.D
 Acq On : 05 Dec 2023 16:20
 Operator : SY/MD
 Sample : 05599-05
 Misc : 4.73g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 LKWD-BACKFILL-03

Quant Time: Dec 06 02:26:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 05 01:24:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.813	168	132617	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.709	114	220784	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.508	117	216419	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	103023	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	87628	71.490	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	142.980%
35) Dibromofluoromethane	7.740	113	70938	53.390	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.780%
50) Toluene-d8	10.197	98	254791	50.696	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.400%
62) 4-Bromofluorobenzene	12.495	95	92549	55.997	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	112.000%
Target Compounds						
17) Carbon Disulfide	4.216	76	3446	1.043	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120523\
Data File : VY016599.D
Acq On : 05 Dec 2023 16:20
Operator : SY/MD
Sample : 05599-05
Misc : 4.73g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
LKWD-BACKFILL-03

Quant Time: Dec 06 02:26:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120423S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 05 01:24:25 2023
Response via : Initial Calibration

