

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120523\
 Data File : VY016601.D
 Acq On : 05 Dec 2023 17:06
 Operator : SY/MD
 Sample : 05599-09
 Misc : 4.78g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 LKWD-BACKFILL-05

Quant Time: Dec 06 02:27:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 05 01:24:25 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	133225	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.709	114	219453	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.501	117	218074	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	103412	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.160	65	86312	70.095	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 140.200%	
35) Dibromofluoromethane	7.734	113	71220	53.927	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.860%	
50) Toluene-d8	10.191	98	255317	51.109	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 102.220%	
62) 4-Bromofluorobenzene	12.489	95	92783	56.479	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 112.960%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.740	84	13226	7.932	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120523\
Data File : VY016601.D
Acq On : 05 Dec 2023 17:06
Operator : SY/MD
Sample : 05599-09
Misc : 4.78g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
LKWD-BACKFILL-05

Quant Time: Dec 06 02:27:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120423S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 05 01:24:25 2023
Response via : Initial Calibration

