

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120523\
 Data File : VY016602.D
 Acq On : 05 Dec 2023 17:30
 Operator : SY/MD
 Sample : 05599-11
 Misc : 4.38g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 LKWD-BACKFILL-06

Quant Time: Dec 06 02:27:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 05 01:24:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	133692	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	219737	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	213081	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	101292	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	86467	69.976	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 139.960%	
35) Dibromofluoromethane	7.734	113	70418	53.251	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.500%	
50) Toluene-d8	10.191	98	251398	50.260	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 100.520%	
62) 4-Bromofluorobenzene	12.495	95	90391	54.952	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 109.900%	
Target Compounds						
17) Carbon Disulfide	4.216	76	4120	1.237	ug/l	Qvalue # 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120523\
Data File : VY016602.D
Acq On : 05 Dec 2023 17:30
Operator : SY/MD
Sample : 05599-11
Misc : 4.38g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
LKWD-BACKFILL-06

Quant Time: Dec 06 02:27:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120423S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 05 01:24:25 2023
Response via : Initial Calibration

