

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120523\
 Data File : VY016604.D
 Acq On : 05 Dec 2023 18:17
 Operator : SY/MD
 Sample : 05599-15
 Misc : 4.59g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 LKWD-BACKFILL-08

Quant Time: Dec 06 02:28:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 05 01:24:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	128869	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	213497	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	208508	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	99779	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	83112	69.778	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 139.560%	
35) Dibromofluoromethane	7.734	113	68398	53.235	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.480%	
50) Toluene-d8	10.191	98	247642	50.956	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.920%	
62) 4-Bromofluorobenzene	12.495	95	89480	55.988	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 111.980%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.735	84	13635	8.692	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120523\
Data File : VY016604.D
Acq On : 05 Dec 2023 18:17
Operator : SY/MD
Sample : 05599-15
Misc : 4.59g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
LKWD-BACKFILL-08

Quant Time: Dec 06 02:28:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120423S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 05 01:24:25 2023
Response via : Initial Calibration

