



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120523\
Data File : VY016587.D
Acq On : 05 Dec 2023 07:52
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120423S.M
Title : SW846 8260
Last Update : Tue Dec 05 01:24:25 2023



AutoFind: Scans 1751, 1752, 1753; Background Corrected with Scan 1743

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.9	16524	PASS
75	95	30	60	53.5	46691	PASS
95	95	100	100	100.0	87224	PASS
96	95	5	9	6.6	5747	PASS
173	174	0.00	2	1.2	925	PASS
174	95	50	100	88.7	77411	PASS
175	174	5	9	7.5	5788	PASS
176	174	95	101	96.9	74979	PASS
177	176	5	9	6.8	5110	PASS