

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120621\
 Data File : VY006980.D
 Acq On : 07 Dec 2021 00:16
 Operator : SY/MD
 Sample : M4884-02RE
 Misc : 3.17g/10.0mL/MSVOA_Y/SOIL
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EW9H8RE

Quant Time: Dec 07 04:14:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\SFAMYLM120321SMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 07 04:00:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.685	114	101475	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.490	117	62972	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.422	152	15268	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.247	65	29491	22.565	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	90.280%	
7) Chloroethane-d5	2.765	69	27518	26.841	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	107.360%	
11) 1,1-Dichloroethene-d2	3.851	63	29653	10.914	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	43.640%#	
21) 2-Butanone-d5	6.887	46	54276	107.524	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	215.040%#	
24) Chloroform-d	7.478	84	62847	24.852	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	99.400%	
26) 1,2-Dichloroethane-d4	8.136	65	40317	26.560	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	106.240%	
32) Benzene-d6	8.106	84	98709	29.509	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	118.040%	
36) 1,2-Dichloropropane-d6	9.118	67	40465	38.501	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	154.000%#	
41) Toluene-d8	10.179	98	64101	20.460	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	81.840%	
43) trans-1,3-Dichloroprop...	10.435	79	11072	21.780	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	87.120%	
47) 2-Hexanone-d5	10.788	63	21408	88.009	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	176.020%#	
56) 1,1,2,2-Tetrachloroeth...	12.556	84	31269	35.711	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	142.840%#	
66) 1,2-Dichlorobenzene-d4	13.721	152	8587	17.246	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	69.000%#	
Target Compounds						
					Qvalue	
13) Acetone	3.948	43	5150	11.132	ug/L	83
27) 1,2-Dichloroethane	8.234	62	4092	2.035	ug/L	96
42) Toluene	10.240	91	13160	2.914	ug/L	96
53) m,p-Xylene	11.697	106	2579	1.404	ug/L	89
54) o-Xylene	12.026	106	1385	0.802	ug/L	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120621\
Data File : VY006980.D
Acq On : 07 Dec 2021 00:16
Operator : SY/MD
Sample : M4884-02RE
Misc : 3.17g/10.0mL/MSVOA_Y/SOIL
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EW9H8RE

Quant Time: Dec 07 04:14:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\SFAMYL120321SMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 07 04:00:28 2021
Response via : Initial Calibration

