

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120621\
 Data File : VY006986.D
 Acq On : 07 Dec 2021 02:28
 Operator : SY/MD
 Sample : M4884-21
 Misc : 6.00g/10.0mL/MSVOA_Y/SOIL
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EW9K5

Quant Time: Dec 07 04:15:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\SFAMYLEM120321SMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 07 04:00:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.691	114	166064	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.489	117	113539	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.422	152	42452	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.247	65	34674	16.212	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	64.840%
7) Chloroethane-d5	2.771	69	30964	18.456	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	73.840%
11) 1,1-Dichloroethene-d2	3.851	63	40268	9.056	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	36.240%#
21) 2-Butanone-d5	6.887	46	46034	55.726	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	111.460%
24) Chloroform-d	7.478	84	79231	19.145	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	76.560%
26) 1,2-Dichloroethane-d4	8.142	65	55600	22.382	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	89.520%
32) Benzene-d6	8.112	84	151009	25.038	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	100.160%
36) 1,2-Dichloropropane-d6	9.118	67	55861	29.479	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	117.920%
41) Toluene-d8	10.179	98	101910	18.041	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	72.160%
43) trans-1,3-Dichloroprop...	10.441	79	19811	21.614	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	86.440%
47) 2-Hexanone-d5	10.788	63	25280	57.641	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	115.280%
56) 1,1,2,2-Tetrachloroeth...	12.556	84	44332	28.081	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	112.320%
66) 1,2-Dichlorobenzene-d4	13.721	152	30029	21.691	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	86.760%
Target Compounds						
					Qvalue	
13) Acetone	3.942	43	2218	2.930	ug/L	94
34) Trichloroethene	8.941	95	186371	99.125	ug/L	93
46) Tetrachloroethene	10.715	164	19651	12.475	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120621\
Data File : VY006986.D
Acq On : 07 Dec 2021 02:28
Operator : SY/MD
Sample : M4884-21
Misc : 6.00g/10.0mL/MSVOA_Y/SOIL
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EW9K5

Quant Time: Dec 07 04:15:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\SFAMYL120321SMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 07 04:00:28 2021
Response via : Initial Calibration

