

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120623\
 Data File : VY016612.D
 Acq On : 06 Dec 2023 13:47
 Operator : SY/MD
 Sample : 05538-01RE
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-02-112223RE

Quant Time: Dec 07 00:32:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 05 01:24:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	35800	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	53727	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.508	117	46546	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.447	152	17092	50.000	ug/l	0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	19920	60.202	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	120.400%
35) Dibromofluoromethane	7.728	113	17435	53.923	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.840%
50) Toluene-d8	10.191	98	60202	49.224	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.440%
62) 4-Bromofluorobenzene	12.495	95	16147	40.147	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	80.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120623\
Data File : VY016612.D
Acq On : 06 Dec 2023 13:47
Operator : SY/MD
Sample : 05538-01RE
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-02-112223RE

Quant Time: Dec 07 00:32:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120423S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 05 01:24:25 2023
Response via : Initial Calibration

