

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120623\
 Data File : VY016613.D
 Acq On : 06 Dec 2023 14:10
 Operator : SY/MD
 Sample : 05599-01RE
 Misc : 4.32g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 LKWD-BACKFILL-01RE

Quant Time: Dec 07 00:33:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 05 01:24:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	126813	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	207588	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.508	117	202754	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	94577	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	81132	69.220	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	138.440%
35) Dibromofluoromethane	7.734	113	65384	52.338	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.680%
50) Toluene-d8	10.197	98	240754	50.949	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.900%
62) 4-Bromofluorobenzene	12.495	95	85458	54.993	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	109.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120623\
Data File : VY016613.D
Acq On : 06 Dec 2023 14:10
Operator : SY/MD
Sample : 05599-01RE
Misc : 4.32g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
LKWD-BACKFILL-01RE

Quant Time: Dec 07 00:33:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120423S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 05 01:24:25 2023
Response via : Initial Calibration

