

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120624\
 Data File : VY020537.D
 Acq On : 06 Dec 2024 15:30
 Operator : SY/MD
 Sample : P5137-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 LAW-OILY-STONES

Quant Time: Dec 07 03:26:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 03 01:39:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	148914	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	255678	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	214242	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	84678	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.066	65	81499	53.813	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.620%
35) Dibromofluoromethane	7.640	113	79778	49.563	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.120%
50) Toluene-d8	10.109	98	308249	50.292	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.580%
62) 4-Bromofluorobenzene	12.407	95	91654	42.756	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	85.520%
Target Compounds						
16) Acetone	3.872	43	55463	147.284	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120624\
Data File : VY020537.D
Acq On : 06 Dec 2024 15:30
Operator : SY/MD
Sample : P5137-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
LAW-OILY-STONES

Quant Time: Dec 07 03:26:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120224S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 03 01:39:23 2024
Response via : Initial Calibration

