

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120624\
 Data File : VY020528.D
 Acq On : 06 Dec 2024 10:54
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Quant Time: Dec 07 03:21:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 03 01:39:23 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 12/09/2024
 Supervised By :Semsettin Yesilyurt 12/09/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	147499	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	227967	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	192490	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	98467	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	69011	46.005	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	92.000%		
35) Dibromofluoromethane	7.646	113	66965	46.660	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	93.320%		
50) Toluene-d8	10.115	98	254853	46.635	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	93.260%		
62) 4-Bromofluorobenzene	12.414	95	88959	46.544	ug/l	0.00
Spiked Amount 50.000	Range 29 - 146		Recovery =	93.080%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.867	85	73664	51.199	ug/l	100
3) Chloromethane	2.074	50	73734	50.615	ug/l	99
4) Vinyl Chloride	2.208	62	75399	50.764	ug/l	99
5) Bromomethane	2.604	94	44945	51.574	ug/l	96
6) Chloroethane	2.745	64	45357	49.824	ug/l	98
7) Trichlorofluoromethane	3.068	101	135790	48.814	ug/l	100
8) Diethyl Ether	3.464	74	37268	47.284	ug/l	94
9) 1,1,2-Trichlorotrifluo...	3.824	101	78399	48.369	ug/l	99
10) Methyl Iodide	4.013	142	93368	51.313	ug/l	95
11) Tert butyl alcohol	4.872	59	24038	247.026	ug/l	100
12) 1,1-Dichloroethene	3.799	96	76054	49.671	ug/l	94
13) Acrolein	3.659	56	17545	192.574	ug/l	99
14) Allyl chloride	4.397	41	136176	49.866	ug/l	97
15) Acrylonitrile	5.067	53	81067	246.853	ug/l	99
16) Acetone	3.872	43	103932	278.642	ug/l	98
17) Carbon Disulfide	4.116	76	231626	52.855	ug/l	100
18) Methyl Acetate	4.391	43	39242	49.500	ug/l	97
19) Methyl tert-butyl Ether	5.122	73	189628	48.495	ug/l	97
20) Methylene Chloride	4.628	84	80415	50.681	ug/l	97
21) trans-1,2-Dichloroethene	5.128	96	85799	50.450	ug/l	95
22) Diisopropyl ether	6.031	45	269396	49.609	ug/l	97
23) Vinyl Acetate	5.970	43	719736	250.579	ug/l	98
24) 1,1-Dichloroethane	5.927	63	163122	49.895	ug/l	99
25) 2-Butanone	6.902	43	119478	253.671	ug/l	99
26) 2,2-Dichloropropane	6.896	77	154299	49.900	ug/l	99
27) cis-1,2-Dichloroethene	6.902	96	99442	50.076	ug/l	98
28) Bromochloromethane	7.256	49	61352	50.253	ug/l	93
29) Tetrahydrofuran	7.268	42	64114	238.611	ug/l	98
30) Chloroform	7.427	83	165256	49.252	ug/l	99
31) Cyclohexane	7.707	56	138651	46.911	ug/l	97
32) 1,1,1-Trichloroethane	7.628	97	157261	48.675	ug/l	98
36) 1,1-Dichloropropene	7.841	75	119832	49.444	ug/l	99
37) Ethyl Acetate	6.994	43	48231	50.232	ug/l	98
38) Carbon Tetrachloride	7.829	117	144677	50.017	ug/l	100
39) Methylcyclohexane	9.115	83	148871	49.502	ug/l	98
40) Benzene	8.085	78	356382	49.874	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120624\
 Data File : VY020528.D
 Acq On : 06 Dec 2024 10:54
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Quant Time: Dec 07 03:21:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 03 01:39:23 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 12/09/2024
 Supervised By :Semsettin Yesilyurt 12/09/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.226	41	26642	48.810	ug/l	92
42) 1,2-Dichloroethane	8.164	62	95762	51.002	ug/l	99
43) Isopropyl Acetate	8.201	43	94393	48.685	ug/l	96
44) Trichloroethene	8.872	130	88865	50.447	ug/l	97
45) 1,2-Dichloropropane	9.146	63	81604	48.802	ug/l	97
46) Dibromomethane	9.237	93	42553	51.398	ug/l	99
47) Bromodichloromethane	9.426	83	122524	50.230	ug/l	96
48) Methyl methacrylate	9.225	41	43910	48.567	ug/l	96
49) 1,4-Dioxane	9.237	88	8531	1078.009	ug/l	96
51) 4-Methyl-2-Pentanone	10.006	43	238163	248.736	ug/l	99
52) Toluene	10.176	92	226650	50.838	ug/l	99
53) t-1,3-Dichloropropene	10.402	75	110537	50.476	ug/l	99
54) cis-1,3-Dichloropropene	9.859	75	132459	51.181	ug/l	96
55) 1,1,2-Trichloroethane	10.579	97	55189	49.972	ug/l	99
56) Ethyl methacrylate	10.444	69	78793	50.422	ug/l	95
57) 1,3-Dichloropropane	10.725	76	99728	50.427	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.713	63	199974	273.318	ug/l	97
59) 2-Hexanone	10.761	43	175974	254.760	ug/l	97
60) Dibromochloromethane	10.914	129	78637	51.143	ug/l	100
61) 1,2-Dibromoethane	11.024	107	48083	47.609	ug/l	95
64) Tetrachloroethene	10.652	164	77566	47.689	ug/l	99
65) Chlorobenzene	11.450	112	235505	48.811	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.524	131	81926	47.879	ug/l	98
67) Ethyl Benzene	11.524	91	445589	49.270	ug/l	99
68) m/p-Xylenes	11.633	106	327409	98.868	ug/l	99
69) o-Xylene	11.962	106	152263	49.079	ug/l	97
70) Styrene	11.975	104	258061	49.857	ug/l	99
71) Bromoform	12.139	173	43237	49.699	ug/l #	99
73) Isopropylbenzene	12.261	105	423521	44.454	ug/l	99
74) N-amyl acetate	12.078	43	81848	43.832	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.511	83	59212	43.218	ug/l	99
76) 1,2,3-Trichloropropane	12.560	75	45009m	44.871	ug/l	
77) Bromobenzene	12.536	156	87883	44.543	ug/l	97
78) n-propylbenzene	12.603	91	504021	44.683	ug/l	99
79) 2-Chlorotoluene	12.688	91	279887	43.973	ug/l	99
80) 1,3,5-Trimethylbenzene	12.743	105	342297	44.834	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.310	75	21620	44.225	ug/l	95
82) 4-Chlorotoluene	12.786	91	289442	44.561	ug/l	99
83) tert-Butylbenzene	13.005	119	310017	44.913	ug/l	99
84) 1,2,4-Trimethylbenzene	13.048	105	336656	45.367	ug/l	99
85) sec-Butylbenzene	13.182	105	449236	45.412	ug/l	100
86) p-Isopropyltoluene	13.298	119	374460	45.807	ug/l	99
87) 1,3-Dichlorobenzene	13.298	146	172673	45.380	ug/l	99
88) 1,4-Dichlorobenzene	13.377	146	166671	44.744	ug/l	99
89) n-Butylbenzene	13.627	91	349933	46.154	ug/l	99
90) Hexachloroethane	13.889	117	72337	45.406	ug/l	94
91) 1,2-Dichlorobenzene	13.663	146	148484	46.016	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.279	75	9136	44.047	ug/l	98
93) 1,2,4-Trichlorobenzene	14.931	180	86224	47.521	ug/l	100
94) Hexachlorobutadiene	15.035	225	60470	47.639	ug/l	99
95) Naphthalene	15.151	128	140403	47.818	ug/l	100
96) 1,2,3-Trichlorobenzene	15.340	180	71260	48.672	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120624\
Data File : VY020528.D
Acq On : 06 Dec 2024 10:54
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024

Quant Time: Dec 07 03:21:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120224S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 03 01:39:23 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024

Abundance

TIC: VY020528.D\data.ms

Time-->

1350000

1300000

1250000

1200000

1150000

1100000

1050000

1000000

950000

900000

850000

800000

750000

700000

650000

600000

550000

500000

450000

400000

350000

300000

250000

200000

150000

100000

50000

0

0.00

0.05

0.10

0.15

0.20

0.25

0.30

0.35

0.40

0.45

0.50

0.55

0.60

0.65

0.70

0.75

0.80

0.85

0.90

0.95

1.00

1.05

1.10

1.15

1.20

1.25

1.30

1.35

1.40

1.45

1.50

1.55

1.60

1.65

1.70

1.75

1.80

1.85

1.90

1.95

2.00

2.05

2.10

2.15

2.20

2.25

2.30

2.35

2.40

2.45

2.50

2.55

2.60

2.65

2.70

2.75

2.80

2.85

2.90

2.95

3.00

3.05

3.10

3.15

3.20

3.25

3.30

3.35

3.40

3.45

3.50

3.55

3.60

3.65

3.70

3.75

3.80

3.85

3.90

3.95

4.00

4.05

4.10

4.15

4.20

4.25

4.30

4.35

4.40

4.45

4.50

4.55

4.60

4.65

4.70

4.75

4.80

4.85

4.90

4.95

5.00

5.05

5.10

5.15

5.20

5.25

5.30

5.35

5.40

5.45

5.50

5.55

5.60

5.65

5.70

5.75

5.80

5.85

5.90

5.95

6.00

6.05

6.10

6.15

6.20

6.25

6.30

6.35

6.40

6.45

6.50

6.55

6.60

6.65

6.70

6.75

6.80

6.85

6.90

6.95

7.00

7.05

7.10

7.15

7.20

7.25

7.30

7.35

7.40

7.45

7.50

7.55

7.60

7.65

7.70

7.75

7.80

7.85

7.90

7.95

8.00

8.05

8.10

8.15

8.20

8.25

8.30

8.35

8.40

8.45

8.50

8.55

8.60

8.65

8.70

8.75

8.80

8.85

8.90

8.95

9.00

9.05

9.10

9.15

9.20

9.25

9.30

9.35

9.40

9.45

9.50

9.55

9.60

9.65

9.70

9.75

9.80

9.85

9.90

9.95

10.00

10.05

10.10

10.15

10.20

10.25

10.30

10.35

10.40

10.45

10.50

10.55

10.60

10.65

10.70

10.75

10.80

10.85

10.90

10.95

11.00

11.05

11.10

11.15

11.20

11.25

11.30

11.35

11.40

11.45

11.50

11.55

11.60

11.65

11.70

11.75

11.80

11.85

11.90

11.95

12.00

12.05

12.10

12.15

12.20

12.25

12.30

12.35

12.40

12.45

12.50

12.55

12.60

12.65

12.70

12.75

12.80

12.85

12.90

12.95

13.00

13.05

13.10

13.15

13.20

13.25

13.30

13.35

13.40

13.45

13.50

13.55

13.60

13.65

13.70

13.75

13.80

13.85

13.90

13.95

14.00

14.05

14.10

14.15

14.20

14.25

14.30

14.35

14.40

14.45

14.50

14.55

14.60

14.65

14.70

14.75

14.80

14.85

14.90

14.95

15.00

15.05

15.10

15.15

15.20

15.25

15.30

15.35

15.40

15.45

15.50

15.55

15.60

15.65

15.70

15.75

15.80

15.85

15.90

15.95

16.00

16.05

16.10

16.15

16.20

16.25

16.30

16.35


16.40

16.45

16.50


16.55

1



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.713 min Scan# 98
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:168 Resp: 147499
Ion Ratio Lower Upper
168 100
99 60.4 46.6 69.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#2
Dichlorodifluoromethane
Concen: 51.199 ug/l
RT: 1.867 min Scan# 24
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54


Tgt Ion: 85 Resp: 73664
Ion Ratio Lower Upper
85 100
87 31.9 16.1 48.2





#3
Chloromethane
Concen: 50.615 ug/l
RT: 2.074 min Scan# 58
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 50 Resp: 73734
Ion Ratio Lower Upper
50 100
52 32.8 26.9 40.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#4
Vinyl Chloride
Concen: 50.764 ug/l
RT: 2.208 min Scan# 80
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54


Tgt Ion: 62 Resp: 75399
Ion Ratio Lower Upper
62 100
64 31.4 24.7 37.1





#5
Bromomethane
Concen: 51.574 ug/l
RT: 2.604 min Scan# 146
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Instrument : MSVOA_Y
ClientSampleId : VSTDCCC050



Tgt Ion: 94 Resp: 44945
Ion Ratio Lower Upper
94 100
96 90.5 75.4 113.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#6
Chloroethane
Concen: 49.824 ug/l
RT: 2.745 min Scan# 168
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Tgt Ion: 64 Resp: 45357
Ion Ratio Lower Upper
64 100
66 33.1 25.7 38.5






Trichlorofluoromethane
 Concen: 48.814 ug/l
 RT: 3.068 min Scan# 222
 Delta R.T. -0.006 min
 Lab File: VY020528.D
 Acq: 06 Dec 2024 10:54

Instrument :

MSVOA_Y

ClientSampleId :




VSTDCCC050



Tgt Ion: 101 Resp: 135796
 Ion Ratio Lower Upper
 101 100
 103 64.5 51.5 77.3




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
 Supervised By :Semsettin Yesilyurt 12/09/2024



Diethyl Ether
 Concen: 47.284 ug/l
 RT: 3.464 min Scan# 286
 Delta R.T. -0.006 min
 Lab File: VY020528.D
 Acq: 06 Dec 2024 10:54


Tgt Ion: 74 Resp: 37268
 Ion Ratio Lower Upper
 74 100
 45 101.3 53.6 160.8





#9
1,1,2-Trichlorotrifluoroethane
Concen: 48.369 ug/l
RT: 3.824 min Scan# 346
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:101 Resp: 78399
Ion Ratio Lower Upper
101 100
85 47.0 38.0 57.0
151 75.9 59.6 89.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024




#10
Methyl Iodide
Concen: 51.313 ug/l
RT: 4.013 min Scan# 376
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Tgt Ion:142 Resp: 93368
Ion Ratio Lower Upper
142 100
127 53.4 38.9 58.3
141 14.6 11.7 17.5





#11
Tert butyl alcohol
Concen: 247.026 ug/l
RT: 4.872 min Scan# 511
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 59 Resp: 24038
Ion Ratio Lower Upper
59 100
57 10.1 8.2 12.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024




#12
1,1-Dichloroethene
Concen: 49.671 ug/l
RT: 3.799 min Scan# 341
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Tgt Ion: 96 Resp: 76054
Ion Ratio Lower Upper
96 100
61 168.7 143.3 214.9
98 62.5 49.8 74.8





#13
Acrolein
Concen: 192.574 ug/l
RT: 3.659 min Scan# 319
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 56 Resp: 17545
Ion Ratio Lower Upper
56 100
55 73.2 57.7 86.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#14
Allyl chloride
Concen: 49.866 ug/l
RT: 4.397 min Scan# 439
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54


Tgt Ion: 41 Resp: 136176
Ion Ratio Lower Upper
41 100
39 69.7 58.8 88.2
76 37.1 29.7 44.5





#15
Acrylonitrile
Concen: 246.853 ug/l
RT: 5.067 min Scan# 541
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 53	Resp: 81067
Ion Ratio	Lower Upper
53	100
52	82.4 65.2 97.8
51	37.8 30.2 45.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#16
Acetone
Concen: 278.642 ug/l
RT: 3.872 min Scan# 353
Delta R.T. -0.012 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54


Tgt Ion: 43	Resp: 103932
Ion Ratio	Lower Upper
43	100
58	30.8 25.7 38.5





#17
Carbon Disulfide
Concen: 52.855 ug/l
RT: 4.116 min Scan# 39
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 76 Resp: 231626
Ion Ratio Lower Upper
76 100
78 9.3 7.3 10.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024




#18
Methyl Acetate
Concen: 49.500 ug/l
RT: 4.391 min Scan# 438
Delta R.T. -0.012 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54



Tgt Ion: 43 Resp: 39242
Ion Ratio Lower Upper
43 100
74 24.4 18.6 27.8









#19
Methyl tert-butyl Ether
Concen: 48.495 ug/l
RT: 5.122 min Scan# 55
Delta R.T. -0.012 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 73 Resp: 189628
Ion Ratio Lower Upper
73 100
57 21.1 18.0 27.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#20
Methylene Chloride
Concen: 50.681 ug/l
RT: 4.628 min Scan# 477
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54


Tgt Ion: 84 Resp: 80415
Ion Ratio Lower Upper
84 100
49 127.2 104.7 157.1
51 37.5 33.4 50.0
86 64.1 52.2 78.2





#21
trans-1,2-Dichloroethene
Concen: 50.450 ug/l
RT: 5.128 min Scan# 559
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 96 Resp: 85799
Ion Ratio Lower Upper
96 100
61 144.6 111.6 167.4
98 63.6 46.7 70.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#22
Diisopropyl ether
Concen: 49.609 ug/l
RT: 6.031 min Scan# 707
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Tgt Ion: 45 Resp: 269396
Ion Ratio Lower Upper
45 100
43 54.2 44.8 67.2
87 28.1 20.6 30.8
59 11.0 9.2 13.8





#23

Vinyl Acetate

Concen: 250.579 ug/l

RT: 5.970 min Scan# 69

Delta R.T. -0.006 min

Lab File: VY020528.D

Acq: 06 Dec 2024 10:54

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 719736

Ion Ratio Lower Upper

43 100





86 10.8 8.0 12.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 12/09/2024

Supervised By :Semsettin Yesilyurt 12/09/2024



#24

1,1-Dichloroethane

Concen: 49.895 ug/l

RT: 5.927 min Scan# 690

Delta R.T. -0.006 min

Lab File: VY020528.D

Acq: 06 Dec 2024 10:54




Tgt Ion: 63 Resp: 163122


Ion Ratio Lower Upper

63 100

98 6.8 3.4 10.1


100 3.5 2.0 5.8





#25
2-Butanone
Concen: 253.671 ug/l
RT: 6.902 min Scan# 851
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 119478
Ion Ratio Lower Upper
43 100
72 26.2 20.4 30.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#26
2,2-Dichloropropane
Concen: 49.900 ug/l
RT: 6.896 min Scan# 849
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54


Tgt Ion: 77 Resp: 154299
Ion Ratio Lower Upper
77 100
97 21.5 10.9 32.7





#27
cis-1,2-Dichloroethene
Concen: 50.076 ug/l
RT: 6.902 min Scan# 850
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion	Ratio	Lower	Upper
96	100		
61	153.9	0.0	312.4
98	64.4	0.0	126.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#28
Bromochloromethane
Concen: 50.253 ug/l
RT: 7.256 min Scan# 908
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Tgt Ion	Ratio	Lower	Upper
49	100		
129	1.3	0.0	4.0
130	74.9	55.0	82.6









#29
Tetrahydrofuran
Concen: 238.611 ug/l
RT: 7.268 min Scan# 911
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 42 Resp: 64114
Ion Ratio Lower Upper
42 100
72 42.7 33.4 50.0
71 41.8 31.9 47.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#30
Chloroform
Concen: 49.252 ug/l
RT: 7.427 min Scan# 936
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54


Tgt Ion: 83 Resp: 165256
Ion Ratio Lower Upper
83 100
85 62.3 50.3 75.5





#31
Cyclohexane
Concen: 46.911 ug/l
RT: 7.707 min Scan# 98
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 56 Resp: 138651
Ion Ratio Lower Upper
56 100
69 33.9 25.3 37.9
84 86.0 66.9 100.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#32
1,1,1-Trichloroethane
Concen: 48.675 ug/l
RT: 7.628 min Scan# 969
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Tgt Ion: 97 Resp: 157261
Ion Ratio Lower Upper
97 100
99 64.5 51.8 77.6
61 43.2 37.1 55.7









#33
 1,2-Dichloroethane-d4
 Concen: 46.005 ug/l
 RT: 8.067 min Scan# 1041
 Delta R.T. -0.006 min
 Lab File: VY020528.D
 Acq: 06 Dec 2024 10:54

Instrument : MSVOA_Y
 ClientSampleId : VSTDCCC050




Tgt Ion:	65	Resp:	69011
Ion	Ratio	Lower	Upper
65	100		
67	50.5	0.0	105.8


Manual Integrations
APPROVED
 Reviewed By :Mahesh Dadoda 12/09/2024
 Supervised By :Semsettin Yesilyurt 12/09/2024



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.622 min Scan# 1132
 Delta R.T. -0.000 min
 Lab File: VY020528.D
 Acq: 06 Dec 2024 10:54

Tgt Ion:	114	Resp:	227967
Ion	Ratio	Lower	Upper
114	100		
63	20.2	0.0	44.8
88	16.5	0.0	32.0









#35
Dibromofluoromethane
Concen: 46.660 ug/l
RT: 7.646 min Scan# 97
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:113 Resp: 66965
Ion Ratio Lower Upper
113 100
111 101.4 81.4 122.0
192 18.7 15.1 22.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#36
1,1-Dichloropropene
Concen: 49.444 ug/l
RT: 7.841 min Scan# 1004
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54


Tgt Ion: 75 Resp: 119832
Ion Ratio Lower Upper
75 100
110 34.5 16.7 50.0
77 31.1 24.6 36.8





#37
Ethyl Acetate
Concen: 50.232 ug/l
RT: 6.994 min Scan# 866
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 48231
Ion Ratio Lower Upper
43 100
61 14.2 10.7 16.1
70 9.9 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#38
Carbon Tetrachloride
Concen: 50.017 ug/l
RT: 7.829 min Scan# 1002
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Tgt Ion:117 Resp: 144677
Ion Ratio Lower Upper
117 100
119 97.1 77.8 116.6
121 31.6 24.6 37.0









#39
Methylcyclohexane
Concen: 49.502 ug/l
RT: 9.115 min Scan# 1213
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 83 Resp: 148871
Ion Ratio Lower Upper
83 100
55 77.1 62.6 93.8
98 47.9 37.0 55.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#40
Benzene
Concen: 49.874 ug/l
RT: 8.085 min Scan# 1044
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54


Tgt Ion: 78 Resp: 356382
Ion Ratio Lower Upper
78 100
77 23.2 19.4 29.0





#41
Methacrylonitrile
Concen: 48.810 ug/l
RT: 7.226 min Scan# 90
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 41 Resp: 26642
Ion Ratio Lower Upper
41 100
39 51.4 49.0 73.4
67 75.5 57.1 85.7
52 33.1 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#42
1,2-Dichloroethane
Concen: 51.002 ug/l
RT: 8.164 min Scan# 1057
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54


Tgt Ion: 62 Resp: 95762
Ion Ratio Lower Upper
62 100
98 8.6 0.0 18.0





#43
Isopropyl Acetate
Concen: 48.685 ug/l
RT: 8.201 min Scan# 1063
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Instrument : MSVOA_Y
ClientSampleId : VSTDCCC050



Tgt Ion: 43 Resp: 94393
Ion Ratio Lower Upper
43 100
61 30.9 22.8 34.2
87 13.6 9.6 14.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#44
Trichloroethene
Concen: 50.447 ug/l
RT: 8.872 min Scan# 1173
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54


Tgt Ion:130 Resp: 88865
Ion Ratio Lower Upper
130 100
95 98.9 0.0 204.8








#45
 1,2-Dichloropropane
 Concen: 48.802 ug/l
 RT: 9.146 min Scan# 1218
 Delta R.T. -0.000 min
 Lab File: VY020528.D
 Acq: 06 Dec 2024 10:54

Instrument : MSVOA_Y
 ClientSampleId : VSTDCCC050






Tgt Ion:	63	Resp:	81604
Ion Ratio	Lower	Upper	
63	100		
65	32.2	24.2	36.4


Manual Integrations
APPROVED
 Reviewed By :Mahesh Dadoda 12/09/2024
 Supervised By :Semsettin Yesilyurt 12/09/2024



#46
 Dibromomethane
 Concen: 51.398 ug/l
 RT: 9.237 min Scan# 1233
 Delta R.T. -0.000 min
 Lab File: VY020528.D
 Acq: 06 Dec 2024 10:54


Tgt Ion:	93	Resp:	42553
Ion Ratio	Lower	Upper	
93	100		
95	82.5	65.3	97.9
174	94.0	74.4	111.6





#47
Bromodichloromethane
Concen: 50.230 ug/l
RT: 9.426 min Scan# 1265
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 83 Resp: 122524
Ion Ratio Lower Upper
83 100
85 61.2 51.4 77.2
127 8.9 6.3 9.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#48
Methyl methacrylate
Concen: 48.567 ug/l
RT: 9.225 min Scan# 1231
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54


Tgt Ion: 41 Resp: 43910
Ion Ratio Lower Upper
41 100
69 88.2 66.1 99.1
39 56.1 46.0 69.0





#49
1,4-Dioxane
Concen: 1078.009 ug/l
RT: 9.237 min Scan# 1233
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 88 Resp: 8531
Ion Ratio Lower Upper
88 100
43 36.5 24.6 37.0
58 75.1 59.6 89.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#50
Toluene-d8
Concen: 46.635 ug/l
RT: 10.115 min Scan# 1377
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Tgt Ion: 98 Resp: 254853
Ion Ratio Lower Upper
98 100
100 64.6 51.3 76.9









#51
4-Methyl-2-Pentanone
Concen: 248.736 ug/l
RT: 10.006 min Scan# 1359
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 238163
Ion Ratio Lower Upper
43 100
58 39.5 30.9 46.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#52
Toluene
Concen: 50.838 ug/l
RT: 10.176 min Scan# 1387
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54


Tgt Ion: 92 Resp: 226650
Ion Ratio Lower Upper
92 100
91 173.8 139.7 209.5





#53
t-1,3-Dichloropropene
Concen: 50.476 ug/l
RT: 10.402 min Scan# 1424
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 75 Resp: 110537
Ion Ratio Lower Upper
75 100
77 31.0 25.2 37.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024




#54
cis-1,3-Dichloropropene
Concen: 51.181 ug/l
RT: 9.859 min Scan# 1335
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54



Tgt Ion: 75 Resp: 132459
Ion Ratio Lower Upper
75 100
77 31.1 25.0 37.4
39 47.3 41.0 61.6





#55

1,1,2-Trichloroethane

Concen: 49.972 ug/l

RT: 10.579 min Scan# 1453

Delta R.T. -0.000 min

Lab File: VY020528.D


Acq: 06 Dec 2024 10:54

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050



Tgt Ion: 97 Resp: 55189

Ion Ratio Lower Upper

97 100

83 87.4 70.8 106.2




85 56.1 44.9 67.3

99 62.0 50.6 76.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#56

Ethyl methacrylate

Concen: 50.422 ug/l

RT: 10.444 min Scan# 1431

Delta R.T. -0.000 min

Lab File: VY020528.D

Acq: 06 Dec 2024 10:54




Tgt Ion: 69 Resp: 78793


Ion Ratio Lower Upper

69 100

41 65.8 56.1 84.1


39 33.7 29.7 44.5





#57
1,3-Dichloropropane
Concen: 50.427 ug/l
RT: 10.725 min Scan# 1477
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Instrument : MSVOA_Y
ClientSampleId : VSTDCCC050



Tgt Ion: 76 Resp: 99728
Ion Ratio Lower Upper
76 100
78 32.7 26.0 39.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#58
2-Chloroethyl Vinyl ether
Concen: 273.318 ug/l
RT: 9.713 min Scan# 1311
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54


Tgt Ion: 63 Resp: 199974
Ion Ratio Lower Upper
63 100
106 27.5 21.0 31.4





#59
2-Hexanone
Concen: 254.760 ug/l
RT: 10.761 min Scan# 1483
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 175974
Ion Ratio Lower Upper
43 100
58 54.6 26.4 79.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#60
Dibromochloromethane
Concen: 51.143 ug/l
RT: 10.914 min Scan# 1508
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Tgt Ion:129 Resp: 78637
Ion Ratio Lower Upper
129 100
127 76.8 38.3 114.9









#61
1,2-Dibromoethane
Concen: 47.609 ug/l
RT: 11.024 min Scan# 1526
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:107 Resp: 48083
Ion Ratio Lower Upper
107 100
109 97.9 74.6 111.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#62
4-Bromofluorobenzene
Concen: 46.544 ug/l
RT: 12.414 min Scan# 1754
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Tgt Ion: 95 Resp: 88959
Ion Ratio Lower Upper
95 100
174 81.0 0.0 156.8
176 78.5 0.0 152.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.420 min Scan# 1592

Delta R.T. -0.006 min

Lab File: VY020528.D

Acq: 06 Dec 2024 10:54

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050

Tgt Ion:	117	Resp:	192496
Ion	Ratio	Lower	Upper
117	100		
82	58.3	48.2	72.4
119	32.5	25.8	38.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#64

Tetrachloroethene

Concen: 47.689 ug/l




RT: 10.652 min Scan# 1465


Delta R.T. -0.000 min

Lab File: VY020528.D

Acq: 06 Dec 2024 10:54


Tgt Ion:	164	Resp:	77566
Ion	Ratio	Lower	Upper
164	100		
166	129.5	103.8	155.8
129	101.0	79.0	118.4
131	94.3	77.1	115.7





#65
Chlorobenzene
Concen: 48.811 ug/l
RT: 11.450 min Scan# 1596
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:112 Resp: 235505
Ion Ratio Lower Upper
112 100
114 32.1 25.7 38.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024




#66
1,1,1,2-Tetrachloroethane
Concen: 47.879 ug/l
RT: 11.524 min Scan# 1608
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54



Tgt Ion:131 Resp: 81926
Ion Ratio Lower Upper
131 100
133 96.5 47.8 143.3
119 62.2 32.5 97.5









#67
Ethyl Benzene
Concen: 49.270 ug/l
RT: 11.524 min Scan# 1608
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 91 Resp: 445589
Ion Ratio Lower Upper
91 100
106 30.5 23.9 35.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#68
m/p-Xylenes
Concen: 98.868 ug/l
RT: 11.633 min Scan# 1626
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Tgt Ion:106 Resp: 327409
Ion Ratio Lower Upper
106 100
91 209.0 166.0 249.0









#69
o-Xylene
Concen: 49.079 ug/l
RT: 11.962 min Scan# 1680
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:106 Resp: 152263
Ion Ratio Lower Upper
106 100
91 225.6 110.0 330.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#70
Styrene
Concen: 49.857 ug/l
RT: 11.975 min Scan# 1682
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Tgt Ion:104 Resp: 258061
Ion Ratio Lower Upper
104 100
78 53.0 41.4 62.2
103 55.2 44.2 66.4









#71
Bromoform
Concen: 49.699 ug/l
RT: 12.139 min Scan# 1709
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion	Ratio	Lower	Upper
173	100		
175	48.3	24.3	73.0
254	0.0	0.1	0.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.352 min Scan# 1908
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Tgt Ion	Ratio	Lower	Upper
152	100		
115	59.8	30.0	90.0
150	173.4	0.0	348.2





#73

Isopropylbenzene

Concen: 44.454 ug/l

RT: 12.261 min Scan# 1729

Delta R.T. -0.000 min

Lab File: VY020528.D

Acq: 06 Dec 2024 10:54

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 105 Resp: 423521

Ion Ratio Lower Upper

105 100





120 26.0 12.7 38.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 12/09/2024

Supervised By :Semsettin Yesilyurt 12/09/2024



#74

N-ethyl acetate

Concen: 43.832 ug/l

RT: 12.078 min Scan# 1699

Delta R.T. -0.000 min

Lab File: VY020528.D

Acq: 06 Dec 2024 10:54

Tgt Ion: 43 Resp: 81848




Ion Ratio Lower Upper


43 100

70 43.2 32.9 49.3

55 25.4 19.3 28.9

61 25.5 19.0 28.4









#75
1,1,2,2-Tetrachloroethane
Concen: 43.218 ug/l
RT: 12.511 min Scan# 1770
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion	Ratio	Lower	Upper
83	100		
131	10.6	4.8	14.4
85	64.2	32.0	96.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#76
1,2,3-Trichloropropane
Concen: 44.871 ug/l m
RT: 12.560 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Tgt Ion	Ratio	Lower	Upper
75	100		
77	0.0	0.0	0.0





#77

Bromobenzene

Concen: 44.543 ug/l

RT: 12.536 min Scan# 1774

Delta R.T. -0.006 min

Lab File: VY020528.D

Acq: 06 Dec 2024 10:54

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050

Tgt	Ion:	156	Resp:	87883
Ion	Ratio	Lower	Upper	
156	100			
77	188.8	97.2	291.6	
158	97.7	47.8	143.4	

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#78

n-propylbenzene

Concen: 44.683 ug/l




RT: 12.603 min Scan# 1785


Delta R.T. -0.000 min

Lab File: VY020528.D

Acq: 06 Dec 2024 10:54

Tgt	Ion:	91	Resp:	504021
Ion	Ratio	Lower	Upper	
91	100			
120	22.3	10.9	32.6	









#79
2-Chlorotoluene
Concen: 43.973 ug/l
RT: 12.688 min Scan# 1799
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 91 Resp: 279887
Ion Ratio Lower Upper
91 100
126 33.4 16.4 49.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#80
1,3,5-Trimethylbenzene
Concen: 44.834 ug/l
RT: 12.743 min Scan# 1808
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54


Tgt Ion:105 Resp: 342297
Ion Ratio Lower Upper
105 100
120 47.4 24.3 72.8





#81
trans-1,4-Dichloro-2-butene
Concen: 44.225 ug/l
RT: 12.310 min Scan# 1737
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion	Ratio	Lower	Upper
75	100		
53	96.6	81.1	121.7
89	47.0	35.1	52.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#82
4-Chlorotoluene
Concen: 44.561 ug/l
RT: 12.786 min Scan# 1815
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Tgt Ion	Ratio	Lower	Upper
91	100		
126	32.7	16.6	49.8





#83

tert-Butylbenzene

Concen: 44.913 ug/l

RT: 13.005 min Scan# 1851

Delta R.T. -0.000 min

Lab File: VY020528.D

Acq: 06 Dec 2024 10:54

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion:	119	Resp:	310017
Ion	Ratio	Lower	Upper
119	100		
91	67.2	33.3	99.9
134	26.0	13.0	38.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 12/09/2024

Supervised By :Semsettin Yesilyurt 12/09/2024



#84

1,2,4-Trimethylbenzene

Concen: 45.367 ug/l




RT: 13.048 min Scan# 1858


Delta R.T. -0.000 min

Lab File: VY020528.D

Acq: 06 Dec 2024 10:54

Tgt Ion:	105	Resp:	336656
Ion	Ratio	Lower	Upper
105	100		
120	43.5	22.0	66.0









#85
sec-Butylbenzene
Concen: 45.412 ug/l
RT: 13.182 min Scan# 1880
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:105 Resp: 449236
Ion Ratio Lower Upper
105 100
134 19.8 9.8 29.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#86
p-Isopropyltoluene
Concen: 45.807 ug/l
RT: 13.298 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Tgt Ion:119 Resp: 374460
Ion Ratio Lower Upper
119 100
134 26.0 12.8 38.4
91 25.4 12.3 36.8





#87

1,3-Dichlorobenzene

Concen: 45.380 ug/l

RT: 13.298 min Scan# 1899

Delta R.T. -0.000 min

Lab File: VY020528.D

Acq: 06 Dec 2024 10:54

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:146 Resp: 172673

Ion Ratio Lower Upper

146 100

111 40.6 20.5 61.6





148 64.6 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 12/09/2024

Supervised By :Semsettin Yesilyurt 12/09/2024



#88

1,4-Dichlorobenzene

Concen: 44.744 ug/l

RT: 13.377 min Scan# 1912

Delta R.T. -0.000 min

Lab File: VY020528.D

Acq: 06 Dec 2024 10:54




Tgt Ion:146 Resp: 166671


Ion Ratio Lower Upper

146 100

111 41.0 19.9 59.6

148 64.3 32.0 96.2









#89
n-Butylbenzene
Concen: 46.154 ug/l
RT: 13.627 min Scan# 1953
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:	91	Resp:	349933
Ion	Ratio	Lower	Upper
91	100		
92	52.7	26.5	79.5
134	25.6	12.2	36.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#90
Hexachloroethane
Concen: 45.406 ug/l
RT: 13.889 min Scan# 1996
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54


Tgt Ion:	117	Resp:	72337
Ion	Ratio	Lower	Upper
117	100		
201	79.0	37.1	111.3





#91
1,2-Dichlorobenzene
Concen: 46.016 ug/l
RT: 13.663 min Scan# 1960
Delta R.T. -0.006 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:146 Resp: 148484
Ion Ratio Lower Upper
146 100
111 42.1 20.8 62.5
148 63.1 31.8 95.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 44.047 ug/l
RT: 14.279 min Scan# 2060
Delta R.T. -0.000 min
Lab File: VY020528.D
Acq: 06 Dec 2024 10:54

Tgt Ion: 75 Resp: 9136
Ion Ratio Lower Upper
75 100
155 82.4 39.7 119.1
157 105.1 52.6 157.8





#93

1,2,4-Trichlorobenzene

Concen: 47.521 ug/l

RT: 14.931 min Scan# 2167

Delta R.T. -0.000 min

Lab File: VY020528.D

Acq: 06 Dec 2024 10:54

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:180 Resp: 86224

Ion Ratio Lower Upper

180 100

182 95.2 47.6 142.8





145 31.4 15.8 47.3

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 12/09/2024

Supervised By :Semsettin Yesilyurt 12/09/2024



#94

Hexachlorobutadiene

Concen: 47.639 ug/l

RT: 15.035 min Scan# 2184

Delta R.T. 0.006 min

Lab File: VY020528.D

Acq: 06 Dec 2024 10:54




Tgt Ion:225 Resp: 60470


Ion Ratio Lower Upper

225 100

223 62.6 31.9 95.5

227 63.8 32.3 96.8





#95

Naphthalene

Concen: 47.818 ug/l

RT: 15.151 min Scan# 2203

Delta R.T. -0.000 min

Lab File: VY020528.D

Acq: 06 Dec 2024 10:54

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050

Tgt Ion:	128	Resp:	140403
Ion	Ratio	Lower	Upper
128	100		
127	13.2	10.5	15.7
129	10.8	8.6	13.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 12/09/2024
Supervised By :Semsettin Yesilyurt 12/09/2024



#96

1,2,3-Trichlorobenzene

Concen: 48.672 ug/l

RT: 15.340 min Scan# 2234

Delta R.T. -0.000 min

Lab File: VY020528.D

Acq: 06 Dec 2024 10:54

Tgt Ion:	180	Resp:	71260
Ion	Ratio	Lower	Upper
180	100		
182	94.0	47.4	142.3
145	32.7	16.9	50.6

