

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120723\
 Data File : VY016627.D
 Acq On : 07 Dec 2023 16:28
 Operator : SY/MD
 Sample : 05617-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HD-02-120423

Quant Time: Dec 07 16:40:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 05 01:24:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.813	168	102307	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.710	114	159504	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.508	117	145050	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.446	152	56947	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	55846	59.059	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.120%
35) Dibromofluoromethane	7.740	113	47385	49.365	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.720%
50) Toluene-d8	10.197	98	177860	48.986	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.980%
62) 4-Bromofluorobenzene	12.502	95	56076	46.964	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	93.920%
Target Compounds						Qvalue
20) Methylene Chloride	4.753	84	8121	5.619	ug/l	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120723\
Data File : VY016627.D
Acq On : 07 Dec 2023 16:28
Operator : SY/MD
Sample : 05617-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HD-02-120423

Quant Time: Dec 07 16:40:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120423S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 05 01:24:25 2023
Response via : Initial Calibration

