

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120723\
 Data File : VY016629.D
 Acq On : 07 Dec 2023 17:14
 Operator : SY/MD
 Sample : 05630-01
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-01-120523

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 12/08/2023
 Supervised By :Semsettin Yesilyurt 12/08/2023

Quant Time: Dec 07 17:57:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 05 01:24:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.820	168	18265	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.716	114	24928	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.514	117	20296	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.447	152	8418	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.173	65	8107	48.022	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.040%
35) Dibromofluoromethane	7.746	113	7611	50.734	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.460%
50) Toluene-d8	10.203	98	26744m	47.130	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.260%
62) 4-Bromofluorobenzene	12.502	95	7569	40.561	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	81.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120723\
Data File : VY016629.D
Acq On : 07 Dec 2023 17:14
Operator : SY/MD
Sample : 05630-01
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-01-120523

Quant Time: Dec 07 17:57:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120423S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 05 01:24:25 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 12/08/2023
Supervised By :Semsettin Yesilyurt 12/08/2023

