

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120723\
 Data File : VY016624.D
 Acq On : 07 Dec 2023 15:18
 Operator : SY/MD
 Sample : VY1207SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1207SBL01

Quant Time: Dec 07 15:55:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 05 01:24:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.819	168	175997	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.715	114	252784	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.514	117	216119	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.446	152	104844	50.000	ug/l	0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.173	65	69705	42.851	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	85.700%
35) Dibromofluoromethane	7.746	113	64447	42.364	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	84.720%
50) Toluene-d8	10.203	98	241526	41.974	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	83.940%
62) 4-Bromofluorobenzene	12.501	95	78882	41.686	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	83.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120723\
Data File : VY016624.D
Acq On : 07 Dec 2023 15:18
Operator : SY/MD
Sample : VY1207SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1207SBL01

Quant Time: Dec 07 15:55:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120423S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 05 01:24:25 2023
Response via : Initial Calibration

