

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120723\
 Data File : VY016628.D
 Acq On : 07 Dec 2023 16:51
 Operator : SY/MD
 Sample : 05639-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 E8965C-CC-001

Quant Time: Dec 07 17:59:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 05 01:24:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.819	168	115368	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.716	114	178128	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.514	117	172459	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.446	152	82511	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.167	65	65696	61.611	ug/l	0.01
Spiked Amount	50.000	Range 50 - 163	Recovery	=	123.220%	
35) Dibromofluoromethane	7.746	113	55747	52.004	ug/l	0.01
Spiked Amount	50.000	Range 54 - 147	Recovery	=	104.000%	
50) Toluene-d8	10.203	98	207051	51.063	ug/l	0.01
Spiked Amount	50.000	Range 58 - 134	Recovery	=	102.120%	
62) 4-Bromofluorobenzene	12.501	95	72083	54.058	ug/l	0.01
Spiked Amount	50.000	Range 30 - 143	Recovery	=	108.120%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.759	84	7894	4.345	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120723\
Data File : VY016628.D
Acq On : 07 Dec 2023 16:51
Operator : SY/MD
Sample : 05639-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
E8965C-CC-001

Quant Time: Dec 07 17:59:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120423S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 05 01:24:25 2023
Response via : Initial Calibration

