

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120924\
 Data File : VY020546.D
 Acq On : 09 Dec 2024 10:54
 Operator : SY/MD
 Sample : VY1209SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1209SBL01

Quant Time: Dec 10 03:57:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 03 01:39:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	178939	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.616	114	304419	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	248651	50.000	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.346	152	92050	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	90635	49.804	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.600%
35) Dibromofluoromethane	7.634	113	93551	48.814	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.620%
50) Toluene-d8	10.103	98	361094	49.481	ug/l	-0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.960%
62) 4-Bromofluorobenzene	12.408	95	106254	41.631	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	83.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120924\
Data File : VY020546.D
Acq On : 09 Dec 2024 10:54
Operator : SY/MD
Sample : VY1209SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1209SBL01

Quant Time: Dec 10 03:57:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120224S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 03 01:39:23 2024
Response via : Initial Calibration

