

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121024\
 Data File : VY020567.D
 Acq On : 10 Dec 2024 14:13
 Operator : SY/MD
 Sample : P5175-01RE
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OK-01-12062024RE

Quant Time: Dec 11 00:49:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 03 01:39:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	20162	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.609	114	28458	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.414	117	20266	50.000	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.346	152	5735	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	8823	43.028	ug/l	-0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	86.060%
35) Dibromofluoromethane	7.634	113	8913	49.750	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.500%
50) Toluene-d8	10.103	98	32832	48.127	ug/l	-0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.260%
62) 4-Bromofluorobenzene	12.401	95	7532	31.568	ug/l	-0.01
Spiked Amount	50.000	Range	29 - 146	Recovery	=	63.140%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121024\
Data File : VY020567.D
Acq On : 10 Dec 2024 14:13
Operator : SY/MD
Sample : P5175-01RE
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OK-01-12062024RE

Quant Time: Dec 11 00:49:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120224S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 03 01:39:23 2024
Response via : Initial Calibration

