

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121024\
 Data File : VY020571.D
 Acq On : 10 Dec 2024 16:10
 Operator : SY/MD
 Sample : P5219-03
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MR-CAM-012

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 12/11/2024
 Supervised By :Semsettin Yesilyurt 12/11/2024

Quant Time: Dec 11 00:50:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 03 01:39:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	32778	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.610	114	50464	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.414	117	32975	50.000	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.347	152	6538	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	18319	54.953	ug/l	-0.01
Spiked Amount 50.000	Range 50 - 163		Recovery = 109.900%			
35) Dibromofluoromethane	7.634	113	16583	52.198	ug/l	-0.01
Spiked Amount 50.000	Range 54 - 147		Recovery = 104.400%			
50) Toluene-d8	10.103	98	52914	43.740	ug/l	-0.01
Spiked Amount 50.000	Range 58 - 134		Recovery = 87.480%			
62) 4-Bromofluorobenzene	12.408	95	11301	26.710	ug/l	0.00
Spiked Amount 50.000	Range 29 - 146		Recovery = 53.420%			
Target Compounds						
16) Acetone	3.867	43	3631	43.806	ug/l	95
17) Carbon Disulfide	4.098	76	4145	4.256	ug/l	99
20) Methylene Chloride	4.629	84	764m	2.167	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121024\
Data File : VY020571.D
Acq On : 10 Dec 2024 16:10
Operator : SY/MD
Sample : P5219-03
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MR-CAM-012

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 12/11/2024
Supervised By :Semsettin Yesilyurt 12/11/2024

Quant Time: Dec 11 00:50:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120224S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 03 01:39:23 2024
Response via : Initial Calibration

