

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121024\
 Data File : VY020576.D
 Acq On : 10 Dec 2024 18:06
 Operator : SY/MD
 Sample : P5194-04
 Misc : 4.50g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WP-2

Quant Time: Dec 11 00:53:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 03 01:39:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	154588	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.610	114	267938	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.414	117	230686	50.000	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.347	152	95035	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	90041	57.271	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 114.540%	
35) Dibromofluoromethane	7.634	113	85553	50.719	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.440%	
50) Toluene-d8	10.103	98	325971	50.750	ug/l	-0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.500%	
62) 4-Bromofluorobenzene	12.402	95	101435	45.154	ug/l	-0.01
Spiked Amount	50.000	Range	29 - 146	Recovery	= 90.300%	
Target Compounds						
16) Acetone	3.867	43	5731	14.660	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121024\
Data File : VY020576.D
Acq On : 10 Dec 2024 18:06
Operator : SY/MD
Sample : P5194-04
Misc : 4.50g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WP-2

Quant Time: Dec 11 00:53:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120224S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 03 01:39:23 2024
Response via : Initial Calibration

